

Knowledge Base Article

Product Group: Microlog Analyzer

Product: CMVA series

Version: N/A

Abstract

The older SKF Microlog Analyzer series data collectors only communicate via serial port. If a new computer is being used that does not have a serial port (if the computer only has a USB port), a USB to serial adapter must be used. This article lists some USB to serial adapters that may be used.

Overview

The following USB to serial adapters have been tested by SKF for compatibility with the Microlog Analyzer:

- Belkin F5U109 USB PDA Adapter
- I/O Gear (ATEN) GUC 232A USB to Serial RS-232 Adapter

The following USB to serial adapters have not been tested by SKF but have been reported to work:

- Belkin F5U409 USB PDA Adapter
- Dynex PL-2303 USB to Serial Adapter

For further assistance, please contact the Technical Support Group by phone at 1-800-523-7514 option 8, or by email at TSG-Americas@skf.com.